

Titre: Microdissected tissue vs tissue slices - a comparative study of tumor explant models cultured on-chip and off-chip. Supplément
Title:

Auteurs: Dina Dorigiv, Kayla Simeone, Laudine Communal, Jennifer Kendall-Dupont, Amélie St-Georges-Robillard, Benjamin Péant, Euridice Carmona, Anne-Marie Mes-Masson, & Thomas Gervais
Authors:

Date: 2021

Type: Article de revue / Article

Référence: Dorigiv, D., Simeone, K., Communal, L., Kendall-Dupont, J., St-Georges-Robillard, A., Péant, B., Carmona, E., Mes-Masson, A.-M., & Gervais, T. (2021). Microdissected tissue vs tissue slices - a comparative study of tumor explant models cultured on-chip and off-chip. *Cancers*, 13(16), 4208 (14 pages).
Citation: <https://doi.org/10.3390/cancers13164208>

 Document en libre accès dans PolyPublie
Open Access document in PolyPublie

URL de PolyPublie: <https://publications.polymtl.ca/9372/>
PolyPublie URL:

Version: Matériel supplémentaire / Supplementary material
Révisé par les pairs / Refereed

Conditions d'utilisation: Creative Commons Attribution 4.0 International (CC BY)
Terms of Use:

 Document publié chez l'éditeur officiel
Document issued by the official publisher

Titre de la revue: *Cancers* (vol. 13, no. 16)
Journal Title:

Maison d'édition: MDPI
Publisher:

URL officiel: <https://doi.org/10.3390/cancers13164208>
Official URL:

Mention légale: © 2021 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<https://creativecommons.org/licenses/by/4.0/>).
Legal notice:

Supplementary tables

Microdissected tissue vs tissue slices - A comparative study of two tumor explant models cultured on-chip and off-chip.

Dina Dorriviv^{1,2}, Kayla Simeone^{1,3}, Laudine Communal¹, Jennifer Kendall-Dupont¹, Amélie St-Georges-Robillard^{1,4}, Benjamin Péant¹, Euridice Carmona¹, Anne-Marie Mes-Masson^{1,3} and Thomas Gervais^{1,2,4 *}

Table S1 dimensions of devices

Parameter	value		
Channel height (mm)	MDT chip	0.9	
	Tissue slice chip	Top layer	0.9
		Bottom layer	1.1
Channel width (mm)	MDT chip	1.1	
	Tissue slice chip	Top layer	1.1
		Bottom layer	1.1
Channel length	MDT chip	6	
	Tissue slice chip	Top layer	23
		Bottom layer	18
Tissue diameter (mm)	MDT	0.5	
	Tissue slice	3	
Tissue thickness (mm)	0.35		
Tissue chamber dimension (mm)	MDT well (cubical)	0.7	
	Tissue slice (cylindrical)	Diameter	7
		Height	2.5
	96 well-plate (Corning ®-cylindrical)	Diameter	6.4
		Height	17

Table S2 tissue uptake parameters and diffusion properties

Parameter	value		
Diffusion constant of glucose (cm ² /s)	Tissue	2.7x10 ⁻⁶	
	Medium	9.27x10 ⁻⁵	
Diffusion constant of oxygen (cm ² /s)	Tissue	1.8x10 ⁻⁵	
	Medium	2.6x10 ⁻⁵	
	PDMS	3.4x10 ⁻⁵	
Saturation concentration (mM)	Oxygen	Tissue	1.02
		medium	0.21
		PDMS	1.43
	Glucose	11	
Partition coefficient	PDMS-Medium	0.15	
	Medium-Tissue	4.8	
Maximum cellular uptake rate (mM/S)	Oxygen	2.07	
	Glucose	1.09	
Michaelis-Menten constant (mM)	Oxygen	4.63x10 ⁻³	
	Glucose	4x10 ⁻²	